


PULSE GENERATOR

PERFORMANCE CHECK


Model: *AVP-MV-15-C-PN*

S.N.: *1492*

Date: *AUG 29 83*



*5 VOLTS/DIV
1.0 NSEC/DIV
400 KHz*



*5 VOLTS/DIV
100 PSEC/DIV
400 KHz.*

- a) Output signal amplitude:
0 TO ±5V
- b) Pulse width:
0 TO 4 NSEC
- c) Rise time:
≤ 60 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:
*110/220V
50-60 Hz.*