

PULSE GENERATOR

PERFORMANCE CHECK

Model:


1720-1

S.N.:

1497

Date:

Jul 30 83



a) Output signal amplitude:

25 TO 100 AMPS

b) Pulse width:

27 NSEC

c) Rise time:

NA

d) Fall time:

NA

e) PRF:

0 TO 1 KHZ

f) Jitter, stability:

OK

g) Prime power:

+15 ± 1.0 V

160 mA (MAX)

60 AMPS/DIV

10 NSEC/DIV

1 KHZ