


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVK-A-C-PN-OS-M

S.N.: 1511

Date: DEC 22, 1983.



-) Output signal amplitude:
9 TO $\pm 15V$
-) Pulse width:
4 TO 100 NSEC
-) Rise time:
 ≤ 100 NSEC
-) Fall time:
 ≤ 2 NSEC
-) PRF: SEE FIGS 1,2,
-) Jitter, stability:
OK
-) Prime power:

10 VOLTS/DIV

2.0 NSEC/DIV

100 KHz.

110, 220V
50-60 Hz