

PULSE GENERATOR

PERFORMANCE CHECK

Model:


PK-V-P

S.N.:

1516

Date:

OCT 26 83



10 VOLTS/DIV
1.0 NSEC/DIV
40 KHZ

a) Output signal amplitude:

$\geq +5V$

b) Pulse width:

0 TO 50 NSEC

c) Rise time:

≤ 100 PSEC

d) Fall time:

≤ 2 NSEC

e) PRF:

0 TO 50 KHZ

f) Jitter, stability:

OK

g) Prime power:

$+15 \pm 1.0V$

250 mA (MAX)

[Signature]