


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVD-P.*

S.N.: *1521*

Date: *DEC 19 83*



- a) Output signal amplitude:  
*+10 TO +50 AMPS*
- b) Pulse width:  
*20 ± 5 NS*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 1 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*20 AMPS/DIV  
10 NS/DIV.*

*+15 ± 1.0 V  
200 mA (MAX)*