


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AK-AU-05-TTL-N*

S.N.: *1522*

Date: *OCT 26 1953*



- a) Output signal amplitude:
0 TO -15V
- b) Pulse width:
0 TO 50 nsec
- c) Rise time:
≤ 100 psec
- d) Fall time:
≤ 2 nsec
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:

*1.0 nsec/div
10 volts/div
40 KHz*

*+15 ± 1.0V
250 mA (MAX)
*[Signature]**