


PULSE GENERATOR

PERFORMANCE CHECK

Model: *PVL-2A-P-T-W-PS*

S.N.: *1524*

Date: *NOV 7 83*



- a) Output signal amplitude:
0 TO +170V
- b) Pulse width:
0 TO 400 NSER.
- c) Rise time:
≤ 1 NSER
- d) Fall time:
≤ 2 NSER
- e) PRF:
0 TO 5 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

*100V/DIV.
10 NSER/DIV
2 KHZ*

*110/220V,
50-60 HZ*

** NEEDS ADDITIONAL
EXTERNAL DELAY LINE
RFL 100-400 NSER
RANGE*