


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVH-5-1-C-P*

S.N.: *1526*

Date: *0*



- a) Output signal amplitude: *0 TO +10V*
- b) Pulse width: *≤ 135 PSEC*
- c) Rise time: *NA*
- d) Fall time: *NA*
- e) PRF: *0 TO 1 MHz*
- f) Jitter, stability: *OK*
- g) Prime power: *110/220V*  
*50-60 MHz*

*10 VOLTS/DIV*  
*1.0 NSEC/DIV*  
*500 KHz*

*[Signature]*