


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PM-2-N*
S.N.: *1534*
Date: *SEPT 15 83*



- a) Output signal amplitude:
0 TO -15V.
- b) Pulse width:
0 TO 2 NSER
- c) Rise time:
≤ 100 PSE.
- d) Fall time:
≤ 135 PSE
- e) PRF:
0 TO 25 MHz
- f) Jitter, stability:
OK
- g) Prime power:

1.0 NSER/DIV
10 VOLTS/DIV
4.0 MHz.

+24 ± 1.0V
500mA (MAX)