


PULSE GENERATOR
PERFORMANCE CHECK

Model: AUP - AU - 1 - C - N

S.N.: 1546

Date: DEC 23 83



- a) Output signal amplitude:
0 TO -10V
- b) Pulse width:
0 TO 4 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 1.0 MHz
- f) Jitter, stability:
OK
- g) Prime power:

10 VOLTS/DIV
1.0 NSEC/DIV
100 KHz

110/220V
50-60 Hz

NOTE UNIT SHIPPED
WITH VOLTAGE SELECTOR
CARD IN THE 220-250
VOLT POSITION.

[Signature]