

PULSE GENERATOR

PERFORMANCE CHECK

Model:


AP-19-14V2-05-N-M

S.N.:

1583

Date:

OCT 31 83



a) Output signal amplitude:

0 TO -20V

b) Pulse width:

0 TO 2 NSER

c) Rise time:

< 100 PSEZ

d) Fall time:

< 200 PSEZ.

e) PRF:

0 TO 1 MHz

f) Jitter, stability:

OK.

g) Prime power:

+15 ± 1.0 V

250 mA (MAX)

10 VOLTS/DIV

1.0 NSER/DIV

400 KHz