

PULSE GENERATOR

PERFORMANCE CHECK

Model:


072P-1A-1A2-05-N-M

S.N.:

1585

Date:

OCT 31 83



10 VOLTS/DIV
1.0 nSEC/DIV
400 KHz.

a) Output signal amplitude:

0 TO -20V

b) Pulse width:

0 TO 2 nSEC

c) Rise time:

< 100 pSEC

d) Fall time:

< 200 pSEC

e) PRF:

0 TO 1.0 MHz.

f) Jitter, stability:

OK

g) Prime power:

+15 ± 1.0 V

250 mA (MAX)