


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-N-*

S.N.: *1590*

Date: *NOV 8 83*



- a) Output signal amplitude:
0 TO -40V
- b) Pulse width:
1 TO 100 NS
- c) Rise time:
< 300 PSEZ
- d) Fall time:
< 2 NS
- e) PRF:
SEE FILE #2
- f) Jitter, stability:
0.1%
- g) Prime power:

32 V/DIV

5 NS/DIV

20 KHz

+15 ± 1.0V

250 mA