


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVN-1-C-PN-M*

S.N.: *1614*

Date: *JUN 30 83*



- a) Output signal amplitude:  
*0 to ±5V*
- b) Pulse width:  
*0.2 to 1.0 nsec*
- c) Rise time:  
*≤ 100 psec*
- d) Fall time:  
*≤ 135 psec*
- e) PRF:  
*50 to 250 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*5 V/DIV*  
*1.0 nsec/DIV*

*110/220V*  
*50-60 Hz*

*[Signature]*