


PULSE GENERATOR

PERFORMANCE CHECK

Model: ANH-5-1M1-C-P

S.N.: 1657

Date: FEB 27 84



32 VOLTS/DIV  
1.0 NSEC/DIV  
100 KHZ

- ) Output signal amplitude:  
0 TO +40V
- ) Pulse width:  
≤ 350 PSEC
- ) Rise time:  
NA
- ) Fall time:  
NA
- ) PRF:  
0 TO 1000000
- ) Jitter, stability:  
OK
- ) Prime power:

110/220V  
50-60 HZ