


PULSE GENERATOR
PERFORMANCE CHECK

Model: *MP-AN-15-MA-05-EN-N*
S.N.: *1662*
Date: *FEB 21 84*



- a) Output signal amplitude:
0 TO 5V
- b) Pulse width:
0 TO 4 NSEC
- c) Rise time:
≤ 60 PSEC
- d) Fall time:
≤ 135 PSEC
- e) PRF:
0 TO 7 MHz
- f) Jitter, stability:
OK
- g) Prime power:

5 VOLTS/DIV
1.0 NSEC/DIV
100 KHz.

+15 ± 1.0V
250 mA (MAX)
LLP