


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVO-1

S.N.: 1670

Date: MAR. 5, 1984



- a) Output signal amplitude:
25 TO 100 μ mps
- b) Pulse width:
25 nsec.
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 KHz
- f) Jitter, stability:
OK
- g) Prime power:

26 μ Amps DIU
10 nsec DIU
1 KHz

+15 \pm 1.0 V
200 mA (on 197X)
LD