


PULSE GENERATOR
PERFORMANCE CHECK

Model: *PK-1*

S.N.: *1679*

Date: *NOV 30 83*



- a) Output signal amplitude:
> +5V
- b) Pulse width:
4 TO 100 nSEC
- c) Rise time: (SEE FILE 1)
≤ 100 PSEC
- d) Fall time:
≤ 2 nSEC.
- e) PRF:
SEE FILE 1, FILE 2
- f) Jitter, stability:
OK
- g) Prime power:

5 V OUTS/DIV
1.0 nSEC/DIV
40 kHz

+15 ± 1.0 V
150 mA (MAX)