

PULSE GENERATOR

PERFORMANCE CHECK

Model:


AE-V-P

S.N.:

1708

Date:

MAR 12 84



a) Output signal amplitude:

0 TO +50V

b) Pulse width:

0 TO 100 NS

c) Rise time:

≤ 300 PS

d) Fall time:

≤ 2 NS

e) PRF:

0 TO 20 KHz

f) Jitter, stability:

OK

g) Prime power:

+15 ± 1.0 V

250 mA (max)

32 VOLTS/DIV

5 NS/DIV

10 KHz