


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVH-HV3-N-05*

S.N.: *1721*

Date: *MAR 12 64*



a) Output signal amplitude:

≈ -60V

b) Pulse width:

≤ 1 nSEC

c) Rise time:

NA

d) Fall time:

NA

e) PRF:

0 to 1 MHz

f) Jitter, stability:

OK

g) Prime power:

32 VOLTS/DIV

2.0 nSEC/DIV

40 KHZ

+15 ± 1.0 V

200 mA MAX