


PULSE GENERATOR
PERFORMANCE CHECK

Model: 19V4-3-P-PS

S.N.: 1732

Date: MVTY 8 84



- a) Output signal amplitude:
0 40 + 500 V
- b) Pulse width:
≤ 5 nsec (half power)
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 to 20 kHz
- f) Jitter, stability:
OK
- g) Prime power:

10 nsec/div
320 volts/div
10 kHz

110/220 V
50-60 Hz
[Signature]