


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVP-AV-15-C-N

S.N.: 1739

Date: APR. 10, 1984



5 VOLTS/DIV  
1.0 nSEC/DIV  
100 kHz

- a) Output signal amplitude:  
0 TO -5 VOLTS
- b) Pulse width:  
0 TO 4 nSEC
- c) Rise time:  
≤ 60 PSEC
- d) Fall time:  
≤ 135 PSEC
- e) PRF:  
0 TO 1 MHz
- f) Jitter, stability:  
OK
- g) Prime power:  
110/220 V  
50-60 Hz