


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVP - AV - 1S - C - PN - M - OS

S.N.: 1744

Date: APR. 10, 1984



- a) Output signal amplitude:  
0 to  $\pm 5$  VOLTS
- b) Pulse width:  
0 to 4 NSEC
- c) Rise time:  
 $\leq 60$  PSEC
- d) Fall time:  
 $\leq 135$  PSEC
- e) PRF:  
0 to 1 MHz
- f) Jitter, stability:  
OK
- g) Prime power:

5 VOLTS/DIV  
1.0 NSEC/DIV  
100 KHz

110/220V  
50-60 Hz

*[Signature]*