


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PVK-A-C-P*

S.N.: *1767*

Date: *MAR 22 84*



*10 VOLTS/DIV*  
*2 nSEC/DIV*  
*100 kHz*

- a) Output signal amplitude:  
*0 TO +15V*
- b) Pulse width:  
*4 TO 100 nSEC*
- c) Rise time:  
*≤ 100 pSEC*
- d) Fall time:  
*≤ 2 nSEC*
- e) PRF:  
*SEE FILE 142*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*110/220V*  
*50-60 Hz*

*[Signature]*