


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AKH-AN-PS-PW*

S.N.: *1784*

Date: *MAR 21 84*



- a) Output signal amplitude:
0 TO ±15V
- b) Pulse width:
0 TO 50 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
OK
- g) Prime power:

2 NSEC/DIV
10 VOLTS/DIV
40 KHz.

110/220V
50-60 Hz

[Signature]