


PULSE GENERATOR

PERFORMANCE CHECK


Model: *AVR-B3-PS-05-PN-W-6A-EW*

S.N.: *1798*

Date: *1 MAY 3 1984*



*P OUT*  
*50 VOLTS / DIV*  
*200 NSEC / DIV*  
*4 KHZ*  
*RL = 50 OHMS*



*N. OUT*  
*50 VOLTS / DIV*  
*2.0 USEC / DIV*  
*4 KHZ*  
*RL = 50 OHMS*

- a) Output signal amplitude:  
*0 TO ± 100 VOLTS*
- b) Pulse width:  
*0.1 TO 10.0 USEC*
- c) Rise time:  
*≤ 5 NSEC*
- d) Fall time:  
*≤ 5 NSEC*
- e) PRF:  
*0 TO 10 KHZ*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*110/220 V*  
*50 - 60 HZ*