


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVN-5-1-C-PN-M*

S.N.: *1803*

Date: *APRIL 30 1984*



- a) Output signal amplitude:
0 TO ±5 VOLTS
- b) Pulse width:
0.2 TO 1.0 NSEC
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 100 PSEC
- e) PRF:
50 TO 250 MHz
- f) Jitter, stability:
OK
- g) Prime power:

3.2 VOLTS/DIV *110/220V*
1.0 NSEC/DIV *50-60 Hz*
≈ 200 MHz (PRF)

A handwritten signature in the bottom right corner of the page.