


PULSE GENERATOR

PERFORMANCE CHECK


Model: *ANMR-2-PS-TRF-LW-PN*

S.N.: *1815*

Date: *JUNE 13 1984*



*5 nSEC/DIV
10 VOLTS/DIV
1 MHz
TR, TF HIGH*



*5 nSEC/DIV
10 VOLTS/DIV
1 MHz
TR, TF LOW*

- a) Output signal amplitude:
0 TO ± 20 VOLTS
- b) Pulse width:
20 TO 200 nSEC
- c) Rise time:
0.3 OR 3.0 nSEC
- d) Fall time:
0.3 OR 3.0 nSEC
- e) PRF:
0 TO 5 MHz
- f) Jitter, stability:
OK
- g) Prime power:

*110/220 V
50-60 Hz*