



PULSE GENERATOR
PERFORMANCE CHECK

Model: AVR-A1-C-05-PN
S.N.: 1821
Date: MAY 15 1984



P OUT:
50 VOLTS/DIV
200 NSEC/DIV
2.0 KHZ
50 OHMS

- a) Output signal amplitude:
0 TO ± 200 VOLTS.
- b) Pulse width:
50 NSEC TO 1.0 USEC
(0.5% DUTY CYCLE MAX)
- c) Rise time:
≤ 10 NSEC
- d) Fall time:
≤ 10 NSEC
- e) PRF:
0 TO 100 KHZ
(0.5% DUTY CYCLE MAX)
- f) Jitter, stability:
OK
- g) Prime power:
115/230 V
50-60 HZ.



N OUT:
50 VOLTS/DIV
200 NSEC/DIV
2.0 KHZ
50 OHMS