


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVD-5B-C-IBM-EW-EP-M-P*  
S.N.: *1827*  
Date: *APRIL 16 84*



- a) Output signal amplitude:  
*0 TO +48 AMP*
- b) Pulse width:  
*40 NSER TO 1.0 USER*
- c) Rise time:  
*≤ 20 NSER*
- d) Fall time:  
*≤ 20 NSER*
- e) PRF:  
*0 TO 1 KHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*200 NSER/DIV*  
*20 VOLTS/DIV*  
*R<sub>L</sub> = 0.75 OHMS*

*110/220V*  
*50-60 Hz*

*∴ I<sub>OUT</sub> ≈ 48 AMPS*  
*PW = 1.0 user*