


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVK-A P*
S.N.: *1852*
Date: *1967 10 84*



- a) Output signal amplitude:
0 TO +15V
- b) Pulse width:
4 TO 100 nSEC
- c) Rise time:
≤ 100 pSEC
- d) Fall time:
≤ 2 nSEC
- e) PRF:
500 Pk 1 + 2
- f) Jitter, stability:
OK
- g) Prime power:

10 VOLTS/DIV
1.0 nSEC/DIV
200 KHz

+15 ± 1.0 VOLTS
200 mA (MAX)
[Signature]