


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVO-3B-P*

S.N.: *1853*

Date: *JUNE 26 84*



- a) Output signal amplitude:
+10 TO +30 VPP
- b) Pulse width:
1.5 NSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 20 KHz
- f) Jitter, stability:
OK
- g) Prime power:
+15 ± 1.0 V

20 AMP/DIV

2.0 NSEC/DIV

10 KHz

R_L = 5.1 Ω