


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AP-AU-2-N*

S.N.: *1856*

Date: *MAY 28 84*



*10 VOLTS/DIV  
2 NS/DIV  
10 KHZ.*

- a) Output signal amplitude:  
*0 TO 10V*
- b) Pulse width:  
*5 TO 50 NS*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 135 PSEC*
- e) PRF:  
*0 TO 50 KHZ*
- f) Jitter, stability:  
*OK*
- g) Prime power:  
*+15V ± 1.0*  
*250 mA (MAX)*

*[Signature]*