


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *PK-A-N*  
S.N.: *1857*  
Date: *MAY 28 84*



- a) Output signal amplitude:  
*0 TO -15 VOLTS*
- b) Pulse width:  
*4 TO 100 NSEC*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 2 NSEC*
- e) PRF:  
*SEE PLS 1.2*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*1.0 NSEC/DIV*  
*10 VOLTS/DIV*  
*40 KHZ*

*+15 ± 1.0 V*  
*1.75 mA (MAX)*

*[Signature]*