


PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-U-C-P*
S.N.: *1873*
Date: *JUNE 28 84*



- a) Output signal amplitude:
0 TO +50 V.
- b) Pulse width:
0 TO 100 NSEC
- c) Rise time:
≤ 300 PSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
0 TO 20 KHZ
- f) Jitter, stability:
OK
- g) Prime power:

32 VOLTS/DIV
10 NSEC/DIV
10 KHZ
(R_L = 50 Ω)

110/220 V
50-60 HZ.