


PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVO-P*

S.N.: *1400*

Date: *June 18 1984*



- a) Output signal amplitude:  
*10 TO 50 mV*
- b) Pulse width:  
*20 ± 5 nsec*
- c) Rise time:  
*NA*
- d) Fall time:  
*NA*
- e) PRF:  
*0 TO 1 kHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*20 Amp/DIV  
10 nsec/DIV  
1 kHz*

*+15 ± 1.0 V  
250 mA  
(MAX)  
*[Signature]**