


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVR-BI-C-N-TR

S.N.: 1903

Date: SEPT 24, 1984



- a) Output signal amplitude:  
0 TO - 50 V
- b) Pulse width:  
20 TO 200 nsec
- c) Rise time:  
≤ 300 psec
- d) Fall time:  
≤ 3 nsec
- e) PRF:  
0 TO 100 kHz
- f) Jitter, stability:  
OK
- g) Prime power:

32 VOLTS/DIV.

5 nsec/DIV

50 kHz.

$R_L = 50 \Omega$

110 - 220 V

50 - 60 Hz.