


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVP-AV-1-C-PN-OS-M

S.N.: 1913

Date: JULY 27, 1984



10 VOLTS/DIV
1.0 NSEC/DIV
100 KHz

- a) Output signal amplitude:
0 TO ± 10 VOLTS.
- b) Pulse width:
0 TO 4 NSEC
- c) Rise time:
 ≤ 100 PSEC
- d) Fall time:
 ≤ 135 PSEC
- e) PRF:
0 TO 1.0 MHz.
- f) Jitter, stability:
OK
- g) Prime power:

110/220V,
50-60 Hz.