

PULSE GENERATOR
PERFORMANCE CHECK

Model:


PTI-N

S.N.:

1433

Date:

Jul 10 1984



) Output signal amplitude:

0 TO -40 VOLTS

) Pulse width:

2 TO 100 NSEC

) Rise time:

≤ 500 PSEC

) Fall time:

≤ 2 NSEC

) PRF:

SEE FIG 1, 2

) Jitter, stability:

OK

) Prime power:

32 VOLTS/DIV

2.0 NSEC/DIV

100 KHZ

+15 ± 1.0V

200 mA (MAX)