

PULSE GENERATOR  
PERFORMANCE CHECK

Model: *AVMP-1-C-PN*  
S.N.: *1954*  
Date: *AUG 20 84*



- a) Output signal amplitude:  
*0 TO ± 10 VOLTS*
- b) Pulse width:  
*5 TO 100 NSEC*
- c) Rise time:  
*≤ 100 PSEC*
- d) Fall time:  
*≤ 200 PSEC*
- e) PRF:  
*0 TO 1 MHz*
- f) Jitter, stability:  
*OK*
- g) Prime power:

*110/220 V,  
50-60 Hz*

*10 VOLTS/DIV.  
2.0 NSEC/DIV.  
1.0 MHz*

*Need  
in CAT  
6  
AVMP  
with*