


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVG - 3 - C - P - M - W

S.N.: 1963

Date: OCT. 9/84



- a) Output signal amplitude:
0 TO 500 V.
- b) Pulse width:
5, 7.5, 10, 12 nsec
(HALF power)
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 10 KHz
- f) Jitter, stability:
OK
- g) Prime power:

10 nsec/DIV
320 VOLTS/DIV
(PW #3 POSITION)

110/220V

50-60 Hz