


PULSE GENERATOR

PERFORMANCE CHECK

Model: AVO-5A-C-AFI-P

S.N.: 1973

Date: OCT 15, 1984.



- a) Output signal amplitude:  
0 TO + 20 AMPS TO
- b) Pulse width: 3  $\mu$   
20 TO 100 NSEC
- c) Rise time:  
 $\approx$  5 NSEC
- d) Fall time:  
 $\approx$  5 NSEC
- e) PRF:  
0 TO 20 KHz
- f) Jitter, stability:  
OK
- g) Prime power:

20 NSEC/DIV

32 VOLTS/DIV

$R_L = 3 \Omega$

PRF = 20 KHz

110/220V

50-60 Hz.