


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVM-2-P.

S.N.: 2004

Date: Nov. 5, 1984.



a) Output signal amplitude:

0 TO +15V...

b) Pulse width:

0 TO 2 NSEC

c) Rise time:

$\leq 100$  PSEC

d) Fall time:

$\leq 135$  PSEC

e) PRF:

0 TO 25 MHz

f) Jitter, stability:

OK

g) Prime power:

+24  $\pm$  1.0 V

550 mA (MAX)

10 VOLTS/DIV  
1.0 NSEC/DIV  
5 mA/DIV