

PULSE GENERATOR
PERFORMANCE CHECK

Model: *AVI-V-HUB-C-DIV*

S.N.: *2011*

Date: *SEPT 10 84*

a) Output signal amplitude:

0 to ± 100 V

b) Pulse width:

2 to 100 nsec

c) Rise time:

≤ 200 psec

d) Fall time:

≤ 2 nsec

e) PRF:

0 to 5 kHz


f) Jitter, stability:

OK

g) Prime power:

110/220 Volts

50-60 Hz



100 Volts/div

2 nsec/div

5 kHz