


PULSE GENERATOR

PERFORMANCE CHECK

Model: *AVK-AU-N-C*

S.N.: *2016*

Date: *SEPT 4 1984*



- a) Output signal amplitude:
0 TO -15V
- b) Pulse width:
2 TO 50 NSEC.
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 2 NSEC
- e) PRF:
0 TO 50 KHz
- f) Jitter, stability:
6K
- g) Prime power:

5 NSEC/DIV
10 VOLTS/DIV
50 KHz

110/220V,
50-60 Hz.