


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVMM-1-N

S.N.: 2021

Date: AUG 28 1984



- a) Output signal amplitude:
0 to -5 VOLTS.
- b) Pulse width:
0.2 to 6.0 NSEC.
- c) Rise time:
≤ 100 PSEC
- d) Fall time:
≤ 200 PSEC
- e) PRF:
0 to 50 MHz.
- f) Jitter, stability:
OK
- g) Prime power:

+24 ± 1.0 VOLT
400 mA (MAX)

5 VOLTS/DIV

5 NSEC/DIV.

40 MHz.