


PULSE GENERATOR  
PERFORMANCE CHECK

Model: AVH-HVI-P

S.N.: 2030

Date: SEPT. 20/84



- d) Output signal amplitude:  
0 TO +100V
- e) Pulse width:  
≤ 1 NSER
- f) Rise time:  
NA
- g) Fall time:  
NA
- h) PRF:  
0 TO 100 KHz
- i) Jitter, stability:  
OK
- j) Prime power:

100 VOLTS/DIV

1.0 NSER/DIV

50 KHz

+15 ± 1.0 V

200 mA

(MAX)