


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-P

S.N.: 2056

Date: SEPT 19, 1984



a) Output signal amplitude:

0 TO +3.0 V

b) Pulse width:

≤ 1 NSEC

c) Rise time:

NA

d) Fall time:

NA

e) PRF:

0 TO 1 MHz

f) Jitter, stability:

OK

g) Prime power:

32 VOLTS/DIV
100 NSEC/DIV
500 KHz

+15 ± 10 V

150 mA (MAX)