


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVH-S-P

S.N.: 2069

Date: SEPT 11 1984



- a) Output signal amplitude:
0 TO +15 VOLTS
- b) Pulse width:
≤ 500 PSEC
- c) Rise time:
NA
- d) Fall time:
NA
- e) PRF:
0 TO 1 MHz
- f) Jitter, stability:
OK
- g) Prime power:

10 VOLTS/DIV
100 nSEC/DIV
500 KHz

+15 ± 1.0 V
200 mA (max)
[Signature]