


PULSE GENERATOR
PERFORMANCE CHECK

Model: AVL-2-PS-P

S.N.: 2076

Date: OCT 26/84



≈ 320 VOLTS/DIV.
10 NSEC/DIV
1 KHz.

- a) Output signal amplitude:
0 TO + 350 VOLTS
- b) Pulse width:
0 TO 100 NSEC.
- c) Rise time:
 ≤ 1.5 NSEC
- d) Fall time:
 ≤ 3 NSEC.
- e) PRF:
0 TO 5 KHz.
- f) Jitter, stability:
OK.
- g) Prime power:
110/220 V
50 - 60 Hz.